



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/688,863	10/17/2003	Terho Hoskonen	3003-00041	9447

26753 7590 01/18/2007  
ANDRUS, SCEALES, STARKE & SAWALL, LLP  
100 EAST WISCONSIN AVENUE, SUITE 1100  
MILWAUKEE, WI 53202

EXAMINER
----------

NASSER, ROBERT L

ART UNIT	PAPER NUMBER
----------	--------------

3735

SHORTENED STATUTORY PERIOD OF RESPONSE	MAIL DATE	DELIVERY MODE
3 MONTHS	01/18/2007	PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

## Office Action Summary

Application No.

10/688,863

Applicant(s)

HOSKONEN ET AL

Examiner

Robert L. Nasser

Art Unit

3735

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 16 October 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-32 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-32 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

Art Unit: 3735

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

The following is a quotation of the first paragraph of 35 U.S.C. 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

Claims 1-20 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the written description requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to reasonably convey to one skilled in the relevant art that the inventor(s), at the time the application was filed, had possession of the claimed invention. Claims 1-20 recite that the electrodes are used to measure substances in the tissue. The only disclosure the examiner sees is for the electrode to measure EEG, EMG, and NMT response signals. None of these are "substances ." Hence, the amendment introduces new matter

Claims 1-20 are rejected under 35 U.S.C. 112, first paragraph, as failing to comply with the enablement requirement. The claim(s) contains subject matter which was not described in the specification in such a way as to enable one skilled in the art to which it pertains, or with which it is most nearly connected, to make and/or use the invention. There is no disclosed substance to be measured. As such, one cannot make and use the invention in that it is unclear what is being measured.

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

Art Unit: 3735

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 21-26 are rejected under 35 U.S.C. 102(e) as being anticipated by Sarkela et al 6950698. Sarkela shows a device with a base element (see figures) having an array of electrodes thereon, and a connector 9. The examiner notes that the phrase "for measuring EEG, EMG, and NMT" is an intended use limitation and is met if the art is capable of the use. The examiner notes that there is no structural difference between an electrode for measured EEG or one for measure EMG or NMT response. Hence, the electrodes of Sarkela are for measuring EEG, EMG, and NMT. With respect to claim 22, two of the electrodes can be stimulus electrodes and the electrodes can be positioned in the locations recited, again noting that the exact placement of the electrodes is an intended use limitation. Claim 23 is rejected in that the arrangement of figure 2 for example has 4 electrodes which can function as recited and be positioned as recited. Claim 24 is rejected in that the electrodes can measure NMT response. Claims 25 and 26 are rejected in that the electrodes of Sarkela can meet the function and can be positioned as recited.

Claims 21-26 are rejected under 35 U.S.C. 102(b) as being anticipated by Bennett et al 623472. Bennett shows a device with multiple electrodes positioned on a

Art Unit: 3735

substrate 24, where the electrodes are NMT electrodes and EMG electrodes and are capable of measuring EEG, and where the electrodes connect through a connector.

Claim 22 is rejected in that the electrodes are positionable as claimed. Claims 23 and 24 are rejected in that depending on the size of the user, the electrodes might be positioned as recited. Claims 25 and 26 are rejected in that the electrodes in the different embodiments are capable of the recite functions, and are capable of being positioned as claimed.

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 27-32 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bennett et al in view of Duckert et al 6389312. Duckert further teaches that it is known to measure NMT response with a mechanical piezoelectric sensor. Hence, it would have been obvious to modify Bennett to use a piezoelectric sensors, as it is merely the substitution of one known equivalent sensor for another.

Applicant's arguments filed 10/16/2006 have been fully considered but they are not persuasive.

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

Cram 5772591 and Devlin et al 6394593 shows a similar device with an array of electrodes mounted on a head mounted support.

Kiani et al 6934570 shows a head mounted electrode assembly with an optical sensor 120.

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Robert L. Nasser whose telephone number is 571 272-4731. The examiner can normally be reached on m-f 9-5.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Charles Marmor II can be reached on 571 272-4730. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3735

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

RLN  
January 6, 2007

Robert L. Nasser  
Primary Examiner  
Art Unit 3735



ROBERT L. NASSER  
PRIMARY EXAMINER